



*Embassy of the United States of America
Phnom Penh, Cambodia*

July 14, 2014

General Services Office
#1, Street 96,
Phnom Penh, Cambodia
Tel: 023 728 000
Fax: 023 728 400

Dear Prospective Offerors:

SUBJECT: Solicitation SCB60014Q0010 – Request for Quotation for “The Supply of Laboratory Equipment for the U.S. Embassy Phnom Penh.”

The U.S. Embassy Phnom Penh invites you to submit a quotation for “The Supply of Laboratory Equipment for Blood Transfusion Center.” A detailed list of the equipment is enclosed.

Your quotation must be submitted in a sealed envelope marked " The Supply of Laboratory Equipment for the U.S. Embassy in Phnom Penh," GSO-Procurement Section, # 1, Street 96, Phnom Penh, Cambodia or be sent by email to PhnomPenhProcurement@state.gov on or before **4:00 PM on July 24, 2014**. No quotation will be accepted after this time.

In order for a quotation to be considered, you must also complete and submit the following:

1. SF-18
2. Specifications and brand name/model of the products offered.

Late quotations will not be accepted and the evaluation method is the lowest price, technically acceptable (LPTA) method.

Direct any questions regarding this solicitation to PhnomPenhProcurement@state.gov during regular business hours.

Sincerely,

Laura Quinn
General Services Officer

REQUEST FOR QUOTATIONS <i>(THIS IS NOT AN ORDER)</i>		THIS RFQ [] IS <input checked="" type="checkbox"/> IS NOT A SMALL BUSINESS-SMALL PURCHASE SET-ASIDE (52.219-4)			PAGE 2	OF 1	PAGES 4
1. REQUEST NO. SCB60014Q0010	2. DATE ISSUED July 14, 2014	3. REQUISITION/PURCHASE REQUEST NO.	4. CERT. FOR NAT. DEF. UNDER BDSA REG. 2 AND/OR DMS REG. 1	5. RATING			
5A. ISSUED BY American Embassy Phnom Penh # 1, St. 96, Sangkat Wat Phnom, Phnom Penh			6. DELIVER BY (Date) 4 – 6 weeks after issuing order				
5B. FOR INFORMATION CALL: (Name and telephone no.) (No collect calls)			7. DELIVERY				
NAME Laura Quinn phnompenhprocurement@state.gov		TELEPHONE NUMBER AREA CODE 023 728 000		X FOB DESTINATION OTHER (See Schedule)			
8. TO:			9. DESTINATION				
a. NAME		b. COMPANY		a. NAME OF CONSIGNEE U.S Embassy Phnom Penh			
c. STREET ADDRESS			b. STREET ADDRESS # 1, St. 96, Wat Phnom				
d. CITY		e. STATE		f. ZIP CODE		c. CITY Phnom Penh, Cambodia	
				d. STATE		e. ZIP CODE	
10. PLEASE FURNISH QUOTATIONS TO THE ISSUING OFFICE IN BLOCK 5A ON OR BEFORE CLOSE OF BUSINESS (Date) July 24, 2014		IMPORTANT: This is a request for information, and quotations furnished are not offers. If you are unable to quote, please so indicate on this form and return it to the address in Block 5A. This request does not commit the Government to pay any costs incurred in the preparation of the submission of this quotation or to contract for supplies or services. Supplies are of domestic origin unless otherwise indicated by quoter. Any representations and/or certifications attached to this Request for Quotations must be completed by the quoter					
11. SCHEDULE (Include applicable Federal, State and local taxes)							
ITEM NO. (a)	SUPPLIES/SERVICES (b)		QUANTITY (c)	UNIT (d)	UNIT PRICE (e)	AMOUNT (f)	
1	Hettich-Benchtop centrifuge, Refrigerated or/Equivalent		01	Set			
2	TERUMO-Separation Stand, Plasma Expresser or/Equivalent		01	Set			
3	TERUMO-Sterile Tubing Welder TSCDII or/Equivalent		01	Set			
4	TERUMO-Transfer Bags or/Equivalent		01	Each			
5	Panasonic-Blood Bank Refrigerator or/Equivalent		01	Set			
6	LMB-Thrombocyte climatic chamber CLIMAX 200 and Agitator or/Equivalent		01	Set			
7	Sartorius- Electronic precision balance or/Equivalent		01	Unit			
		(Details specification enclosed)					
12 DISCOUNT FOR PROMPT PAYMENT		a. 10 CALENDAR DAYS %	b. 20 CALENDAR DAYS %	c. 30 CALENDAR DAYS %	d. CALENDAR DAYS NUMBER %		
NOTE: Additional provisions and representations [X] are [] are not attached.							
13 NAME AND ADDRESS OF QUOTER			14 SIGNATURE OF PERSON AUTHORIZED TO SIGN QUOTATION		15 DATE OF QUOTATION		
a. NAME OF QUOTER							
b. STREET ADDRESS			16. SIGNER				
c. COUNTY			a. NAME (Type or print) Laura Quinn			b. TELEPHONE	
d. CITY	e. STATE	f. ZIP CODE	c. TITLE (Type or print) General Services Officer			AREA CODE NUMBER	

Specifications of Laboratory Equipments:

No.	Required Technical Specifications		Set
1	Hettich - Benchtop centrifuge, Refrigerated or/Equivalent model	<ul style="list-style-type: none"> - Power supply: 200-240v - Frequency: 50Hz - Consumption: 1,800 VA - Emission, Immunity: EN/IEC 61326-1, Class B - Max. capacity: 4x750ml - Max. RPM (Speed): 15,000/mn - Max. RCF: 24,400 - Running time: 1 sec-99h:59min:59sec, ∞ continuous run, short cycle mode - Dimensions (HxWxD): 456x770x706mm - Weight: approx. 141kgs - Temperature control, infinitely available: from -20 to +40°C * Accessories: <ul style="list-style-type: none"> - ROTANTA 460R, Benchtop centrifuge, refrigerated 220V : 1 pcs - Swing-out Rotor : 1 pcs - Bucket in rotor : 4 pcs - Lid for suspension : 4 pcs 	01
2	TERUMO-Separation Stand, Plasma Expresser or/ Equivalent model	<ul style="list-style-type: none"> - The stand is used to extract blood components from a blood collection bag, or from a satellite bag to a transfer bag - It can separate not only large quantities of whole blood but also small quantities of platelets - Accommodates bags up to 500ml in size and is easy to operate 	01
3	TERUMO-Sterile Tubing Welder TSCDII or/Equivalent model	<ul style="list-style-type: none"> - Tubing specifications: PVC Tubing (O.D: 3.9 -4.5mm, I.D: 2.9 - 3.1mm) - Wafer: TSCD wafers - Operating conditions: Temperature: +10 to +40°C, Relative humidity: 10 to 80% (with no condensation), Maximum altitude: 2,000m - Storage conditions: Temperature: -10 to +50°C, Relative humidity: 10 - 95% (with no condensation) - Electrical Requirements: 120/230/240V, 50/60Hz - External Dimensions (WxHxD): 224 x 177 x 342mm - Weight: approx. 6.5kgs * Accessories: <ul style="list-style-type: none"> - TSCD Wafer : 1 pcs 	01
4	TERUMO-Transfer Bags or/Equivalent	<ul style="list-style-type: none"> - 300ml transfer bag with one spike 	01
5	Panasonic-Blood Bank Refrigerator or/Equivalent model	<ul style="list-style-type: none"> - Exterior dimension (WxDxH): 800 x 832 x 1810mm - Interior dimension (WxDxH): 640 x 560 x 1240mm - Effective capacity: 425 litres - Storage capacity: 120 bags (450ml) - Exterior cabinet: Galvanized steel with bake-on finish - Interior cabinet: Stainless Steel - Cabinet Insulation: Foamed-in-place rigid polyurethane - Outer door: triple layer glass windows - Door lock: 1 	01

		<ul style="list-style-type: none"> - Inner door (acrylic): 5 - Shelves: 5 coated hard steel wire shelves - Casters: 4 - Compressor: 160W hermetically sealed - Refrigerant: HFC - Temperature: 4°C ± 1°C (Ambient temperature 35°C) - Air circulation: Force air circulation - Defrost: Fully automatic - Temperature control: Microprocessor - Temperature alarm: High (6°C), Low (2°C), Audible and visual alarm - Power failure alarm: Audible and visual alarm (9 hours), Automatic rechargeable battery (Ni-Cd) - Door Alarm: Visual alarm - Lighting: 15 W Fluorescent lamp - Access port: 1 - Accessories: 30-day recorder, Door lock key 1 set - Net weight: 185kgs 	
6	LMB-Thrombocyte climatic chamber CLIMAX 200 and Agitator or/Equivalent model	<ul style="list-style-type: none"> - External dimensions: 121 x 71 x 201.5cm - Internal dimensions: 111 x 44.5 x 46cm (x3 chambers in heights) - Capacity: 6 x Agitator 24/48 or 3 x Agitator 96 - Finished and packed unit: 135x85x222cm, 325kgs - The weight without internal fitting: 230kgs - *Accessories: <ul style="list-style-type: none"> - Controller and printer: 1 pcs - Thrombocyte-shaker 96: 3 pcs 	01
7	Sartorius- Electronic precision balance or/Equivalent model	<ul style="list-style-type: none"> - Capacity: 4200kg - Readability: 0.01g - Tare range (subtractive): 4200g - Response time (avg): 1.1s - Repeatability: <= ±0.01g - Linearity: <= ±0.02g - Pan: 180 x 180mm - Housing (WxDxH): 230x310x91mm - Net weight: approx. 3.6kgs 	01

Special Requirements:

- *Quotation will be evaluated **Item-by-Item**.*
- *Quotation shall be provided the detail technical specification offered by including product brochure or / catalog for each item.*
- **Payment Terms:** *Full payment will be made within 30 days upon receipt of the products and invoices.*
- **Invoice:** *Invoice should be clearly stated Purchase Order (PO) reference number and send to:*
Financial Management Officer
1, St. 96, Sangkat Wat Phnom, Daun Penh
Phnom Penh, Cambodia
Email: PHPVoucher@state.gov